




XIV Reunión IberNAM “Microsistemas y Nanotecnología”


28 y 29 de Septiembre 2017.
Tordesillas, Valladolid



Universidad de Valladolid
Centro «Tordesillas»
de Relaciones con Iberoamérica



Universidad de Valladolid



**DIPUTACIÓN
DE VALLADOLID**



PÓSTERES

P01. Capacitance-to-Voltage Converter Based on a Programmable Transimpedance Amplifier for Capacitive MEMS Accelerometers	Nicolás Jesús Medrano
P02. Optimization of the aptamers' immobilization conditions for maximizing the response of a dual-aptasensor for cancer biomarker detection	Sofia G. Meirinho
P03. A CMOS Low-Power Tunable Band-Pass Filter for PSD Sensor Applications	Nicolás Jesús Medrano
P04. A Low Pass Filter With sub-Hz Cutoff Frequencies for Portable Read-out Sensor Systems	Nicolás Jesús Medrano
P05. MoS ₂ -Carbon Nanotube Hybrid Material for Gas Sensing	Juan Casanova-Cháfer
P06. Detección de NO ₂ a temperatura ambiente con sensores resistivos basados en nanofibras de SnO ₂	José Pedro Santos
P07. Smart sensor for pneumatic combined clutch and brake	Fernando Martínez
P08. Influence of the bottom-electrode configuration in the performance and reliability of micromachined gas sensors based on nanowires	Stella Vallejos
P09. Electrochemical biosensors modified with subphthalocyanines for phenol detection	Rocio González Antón
P10. Sensores nanoestructurados por deposición de diferentes materiales sensibles. Técnica Layer by Layer	Coral Salvo
P11. Sensores voltamétricos basados en la combinación de nanopartículas de oro y ftalocianinas en el análisis de catecol	Ana I. Ruiz Carmuega
P12. Design and fabrication of gold-coated nano-structured silicon wafers for LDI-MS analysis	Stefania Alexandra Iakab
P13. Sensores basados en espumas de carbono para la	Cristina Fernández Blanco



P02. Optimization of the aptamers' immobilization conditions for maximizing the response of a dual-aptasensor for cancer biomarker detection

Sofia G. Meirinho^a, Maha Ezzedine^{b,c}, Ana C.A. Veloso^{d,e}, Luís G. Dias^{b,f}, Lúcia R. Rodrigues^c, Ali Othmane^g, António M. Peres^{a,*}

^aAssociate Laboratory LSRE-LCM, ESA, Instituto Politécnico de Bragança, Bragança, Portugal; ^bESA, Instituto Politécnico de Bragança, Bragança, Portugal; ^cISBM, Université de Monastir, Monastir, Tunisia; ^dInstituto Politécnico de Coimbra, ISEC, DEQB, Coimbra, Portugal; ^eCEB - Centre of Biological Engineering, University of Minho, Braga, Portugal; ^fCQ-VR – Centro de Química – Vila Real, University of Trás-os-Montes, Vila Real, Portugal; ^gLaboratory of Interfaces and Advanced Materials, Faculty of Medicine of Monastir, University of Monastir, Monastir, Tunisia
e-mail: sgmeirinho@ipb.pt, Associate Laboratory LSRE-LCM, ESA, Instituto Politécnico de Bragança, Bragança, Portugal

Osteopontin (OPN) is a protein that is present in several body fluids and has been reported as a possible cancer biomarker, being its overexpression associated with tumour progression and metastasis [1,2]. A simple and sensitive method that allows the simultaneous detection of single or multiple cancer biomarkers is envisaged and may be an important tool in cancer diagnosis. In this work, two bioreceptors specific for OPN, a RNA aptamer (OPN-R3) previously described by Mi and co-workers [3] and a DNA aptamer (C10K2) developed by our research group, were biotinylated and immobilized on a dual-screen printed gold electrode through streptavidin-biotin interaction. The voltammetric signals generated by the dual-aptasensor array, after the formation of the aptamers-protein complex, were monitored using cyclic voltammetry (CV) and square-wave voltammetry (SWV), using $[\text{Fe}(\text{CN})_6]^{3-/4-}$ as a redox probe. The optimal immobilization conditions for the dual-aptasensor array were established by response surface methodology. The maximum voltammetric response was obtained for a 0.5 μM aptamer concentration after 20 min of aptamers' immobilization and 30 min of aptamer-OPN interaction time at an incubation temperature of 4°C. The satisfactory preliminary results obtained, although needing further confirmation for synthetic or real human samples, point out that the proposed electrochemical dual-aptasensor array could be a simple and sensitive tool for the detection of OPN, as well as for other potential cancer biomarkers and therefore, may be applied in the future for cancer disease monitoring.

[1] Briones-Orta, M.A.; Avendaño-Vázquez, S.E.; Aparicio-Bautista, D.I.; Coombes, J.D.; Weber, G.F.; Syn, W.-K., *Biochim Biophys Acta*. **2017**, 1868, 93-108.

[2] Weber, G.F., 2011. *Cancer Genomics & Proteomics*. **2011**, 8, 263–288.

[3] Mi, Z., Guo, H., Russell, M.B.; Liu, Y.; Sullenger, B. A.; Kuo, P.C., *Mol. Ther.* **2009**, 17, 153–161.

This work was also financially supported by Project POCI-01-0145-FEDER-006984 – Associate Laboratory LSRE-LCM funded by FEDER through COMPETE2020 and by national funds through FCT, Portugal. S. Meirinho also acknowledges the research grant provided by Project UID/EQU/50020/2013.